

# Intelligent Cleaning: Value Add for Medical Device Manufacturers through Validation

by David A. Vogel, Ph.D.  
Intertech Engineering Associates, Inc.

as published in CleanTech,  
February, 2005



*Medical device manufacturers have certain obligations when it comes to validating software used in the design, development, manufacture or control of quality of the device. System and instrumentation manufacturers can stand out in their competitive markets by assisting their clients with their validation needs.*

Medical device manufacturers are required by federal regulation to validate any software that is used in their devices, or that is used in the development or manufacture of the device. Suppliers of instruments or subsystems used by medical device manufacturers can create a market advantage for themselves by taking control of validation and assisting their customers with the regulatory requirements.

Medical device manufacturers that use manufacturing or test instruments that rely on software may be required by federal law to validate that software. This may include software that is embedded as part of the instrument, or external, even off-the-shelf, software that interacts with the instrument. This article does not address software that is embedded in a medical device; it relates only to software or software embedded in instruments used by the medical device manufacturer (MDM) in the design, development, manufacture or control of quality of the device.

Often, there is little information available from the instrument manufacturers to assist their users in their need to validate the instruments for their intended use. This creates an opportunity for system and instrumentation manufacturers to stand out in their competitive markets by assisting their users with their validation needs.

Validation costs can easily exceed the purchase cost of the instrument or software. Whatever the suppliers of software-controlled instruments can do to reduce their customers' cost of validation, therefore, reduces the overall cost of use of the

## COMPANY PROFILE

### Intertech Engineering Associates, Inc.

**Address:** 100 Lowder Brook Avenue  
Suite 2500  
Westwood, MA 02090  
www.inea.com - (781) 801-1100

**Industry:** (Electro)Medical Devices

**Services:** Assessments  
Training  
Consulting  
Hands-on Engineering

**Skills:** Product Design  
Risk Management  
Requirements Engineering  
Electronics Development  
Software Development  
Software Verification and Validation  
Production/Quality System Software Validation

system. Today, this can be a marketing advantage. Eventually it will become a necessity in order to stay competitive in the medical device market.

MDMs are bound by a specific federal regulation that refers to software found in manufacturing and quality systems. This regulation, found in the Federal Code section on medical device production and process controls, states: "(i) Automated Processes. When computers or automated data processing systems are used as part of the production or the quality system, the manufacturer shall validate computer software for its intended use according to an established protocol. All software changes shall be validated before approval and issuance. These validation activities and results shall be documented." 21 CFR 820.70(i)

This is not to be confused with the regulations relating to software embedded in the medical devices themselves. This regulation relates to all other software that is used in the development, manufacture, or control of quality of the device.

The FDA has published the *General Principles of Software Validation; Final Guidance for Industry and FDA Staff* - January 11, 2002 (commonly referred to as the GPSV) to expand on the FDA's interpretation of all legislated software-validation requirements relating to medical devices.

### Should I Worry?

MDMs should inventory all of the software that is used in a production, research, engineering, sales, or service facility. That inventory should include the "intended use" for each software item. If the software item is used to monitor or control part of the manufacturing process, design and development process, or any other part of the quality system, it falls within the regulation. Software to be considered includes that which is:

- Used in the design and development of the medical devices - such as CAD systems, compilers, test software, simulators, code generators.
- Used in the manufacture of the medical devices - such as embedded machine tool

software, machine vision quality control, robotics, statistical process control software.

- Used in the creation, maintenance and control of quality data about the medical devices, such as complaint-handling systems, lot-tracking systems, training systems, QC systems, patient-tracking systems.
- Used in the creation, maintenance, reporting, validation, and storage of electronic records and electronic signatures.

Unregulated software can fall under the requirements of 820.70(i) by its association with other regulated software. A software item that is normally considered unregulated under these guidelines might produce data that is used by regulated software. The validation requirements of 820.70(i) may then apply to that formerly unregulated software item. The validation requirements in this case could be focused on the functionality of the (formerly unregulated) software item related to the creation, storage, manipulation and security of that data used by the regulated software.

### Validation Obligations

MDMs have certain obligations for validating any software that is "regulated." Validation involves many activities in addition to testing. Validation, as recommended by the FDA's GPSV, includes the following:

*Determination of the Software Lifecycle* - Identify the software lifecycle to facilitate planning of validation activities at each phase of the lifecycle.

*Document Intended Use* - The intended use of the software item is similar to a product-level or user-level requirement. Note that the software item need only be validated for its intended use.

*Risk Analysis and Risk Management* - MDMs are required to analyze how failure of a software item would impact the medical device itself. Additionally, the safety of the user, effects on the environment, and regulatory exposure should be considered as risks.

Identified unacceptable risks must be managed by designing and implementing risk-control

measures. Risk can be managed by reducing the severity of a hazardous situation, by reducing the probability of occurrence of the situation, or both. Sufficient risk-control measures should be put in place so that the residual risk is reduced to an acceptable level.

*Configuration, Management and Version Control* - MDMs must have configuration-management plans in place for each software item. These plans should take into consideration who will make the decisions about upgrading the software, who will supply the upgrades, who will install the software, and who will take responsibility for revalidating the software before it goes online.

*Planning* - Quality plans and software-verification-and-validation plans detail the activities, roles, responsibilities, resources and deliverables related to an individual software item or the collection of software items that make up an automated process.

*Technical Evaluations and Management Reviews* - Technical evaluations should be conducted to determine if the software is technically up to the intended use before putting it online. Management reviews reference the output of the technical evaluations and also consider if the residual risk of a failing software item is acceptable prior to deployment of the software.

Management also needs to consider whether the users of the software are trained sufficiently to successfully deploy the software. The management review should be considered the final gate to deployment.

*Testing* - Testing by the users of the software should focus on any risk-control measures that were implemented and then on the functionality of the software whose failure could lead to severe consequences. There is little value in testing functions of the software that have little effect on the manufactured device.

For electronic-record systems, focus testing on functions whose failure could result in record loss, alteration or loss of security.

*Traceability* - A documented trace is needed to show that all of the functionality detailed in the

intended-use document was implemented or acquired. Any risk-control measure identified in your risk-management report must be traced to where it was implemented and where its ability to mitigate the risk was tested.

Finally, functional testing of the software should be traceable back to software requirements or intended use functionality.

Some level of interpretation of the validation guidelines is necessary by MDMs. Validation is to be proportional to the risk and the complexity of the software; so it is clear that risk analysis is needed for all software items.

MDMs will need to address each of the components of validation. However, for embedded software that is acquired from an instrument vendor, the MDM has little control or visibility into the software development process. This takes some creativity on the part of the MDM to validate such software.

## Validation Ideas

*Make Product Validation Profitable* - Validating manufacturing and quality-system software is burdensome for MDMs. The typical MDM user responsible for the validation is not a software engineer and is not familiar with software validation activities.

*Often, there is little information available from the instrument manufacturers to assist their users in their need to validate the instruments for their intended use.*

Vendors of software-controlled instruments might profit by selling validation packages to their MDM customers. These packages might include partial validation documentation, training and/or consulting with MDM users on validating for their intended use. Validated or partially validated software for regulated medical-device manufacturer uses will be a marketing advantage.

*Implement Design Controls and Validation Practices* - Instrument manufacturers can alleviate many

of the validation obligations of their customers if they document that they have followed good software engineering practices by implementing design controls and the components of validation under their control. The details of this are well documented in the FDA guidelines.

*Verification Testing* - Good software-engineering development practices depend on detailed requirements and functional specifications. Verifying the implementation of the software involves testing each requirement in the specifications for its existence and proper function. This testing should be documented in detailed test protocols. The instrument supplier cannot validate for the MDM user's intended use, but it can verify that the software correctly implements the requirements specified.

*Encourage Audits* - Detailed requirements specifications, test protocols, test results, evaluation and review minutes dictated by a controlled design environment are likely to be considered proprietary information by instrument manufacturers.

Encourage your MDM customers to audit your design process, making the validation and design control documentation available to them for viewing. This will satisfy some of your MDM customer's obligations for validation.

*Make Known Defect Lists Available* - A well-tested instrument or software item is likely to have a list of residual defects at release whose resolution was deferred for future releases. After release to the field, the users of the software might discover additional defects.

These defect lists should be made available to the users and prospective users of the software or software driven instrumentation so that they can assess the risk of the defect in their intended-use environment.

*Coach Your Customer* - Instrument suppliers cannot take full responsibility for the validation of their embedded software used by MDMs. Only the user of the instrument can define the specific intended use of the instrument. The risk associated with the use of the instrument will depend on the intended

use; so that too must be the responsibility of the MDM user.

On the other hand, the instrument supplier has experience with the instrument and possibly experience with how MDMs are using the instrument. Instrument suppliers can coach their MDM users in identifying and documenting intended uses. Further, instrument suppliers can leverage their knowledge of instrument failure modes to assist their MDM users in risk analysis and in identifying risk-control measures to manage the risk to acceptable levels.

*Partner with Validation Experts* - MDM users of software-driven instruments often are not software professionals and often have no experience with software validation.

Instrumentation suppliers may not have knowledge of the regulated medical device industry and may not have enough business in the medical device industry to develop in-house expertise in dealing with users' validation needs.

An outsourced validation expert can consult with the instrument supplier on improving internal software development processes to survive customer audits. Additionally the validation expert can work with customers to assist them in final validation for their intended use. The instrument supplier benefits because validation no longer seems like an obstacle in the acquisition of the instrument.

### **Case in Point: Intelligent Cleaning System**

Take for example, an intelligent cleaning system that controls clean and rinse times, temperature and delivery of cleaning agents. The system is controlled by a microprocessor and can store multiple cleaning profiles for different parts that are to be cleaned.

A manufacturer of a drug-delivery device wants to purchase two systems to use for two different purposes. One machine will be used to clean circuit boards before they are assembled into the device. The second machine will be used to clean and sterilize the disposable elements that sell with the device.

The cleaning system manufacturer should, as part of good engineering practice and design control, document product requirements, performance requirements and limitations, and specify any communications protocols between the cleaning system and external controllers or monitors.

They should also design verification test protocols for the system, specifically testing each requirement for the software. Tests should be executed and results formally documented. A known defect list should be prepared from the

*Validation testing should be conducted to provide a level of assurance that failure modes do not exist, that performance specifications can be met, and that control measures that are identified are in place.*

results of the verification testing for any defects in the released product. And it's important to make all of the above results available to the customer. Proprietary information can be made available in audits.

The medical device manufacturer is responsible for identifying the intended uses of the cleaning system. The intended use should be specific enough to identify what performance requirements are expected of the cleaning system. Note that in this example, one use might be considered to be non-critical, such as cleaning of circuit boards, and the other use is critical since it controls sterilization of the disposable element. This means considering how changes to the software embedded in the cleaning system will be controlled. For example, what happens if a system controller with Version 1.0 embedded software fails and is replaced with a new motor with Version 2.0 software? Is there a way to know? Who will be responsible for revalidating the new controller with the new software in the manufacturing system? If there is a way to update software with the system in place, how will unauthorized updates be prevented?

Consider how the cleaning system might fail in a

way that the end product will be affected. Assign severity levels to the anticipated medical device defects relative to the ultimate safety impact to the patient. For example, a failure of the system that would allow a circuit board to pass through the system uncleaned would be a much lower severity than if it allowed the disposable element to pass through unsterilized. It is necessary to perform risk management for the more severe and/or probable risks. Determine how the failure modes could be detected or corrected as part of the overall manufacturing system.

Finally, validation testing should be conducted to provide a level of assurance that risky failure modes do not exist, that critical performance specifications can be met, and that any identified risk-control measures that are identified are in place and do reduce the risk to an acceptable level.

This is an abbreviated itemization of tasks for the supplier of a software-controlled instrument, in this case a cleaning system, and the MDM user of the instrument. A competitive cleaning system supplier who does not perform the validation tasks suggested above, puts additional validation burden on the MDM customer. Conversely, a competitive cleaning system supplier who does perform the validation tasks suggested above, and who assists with some of the user's responsibilities by coaching, training or consulting, creates added value and a marketing advantage for his cleaning system.

In some cases, the MDM may be using an instrument for a non-critical purpose in the manufacturing process, or may be using it in a way in the process that 100 percent of all product is inspected and verified downstream in the process. In these situations a case can be made for reduced validation rigor. A supplier that can assist its user in this type of analysis and documentation can save the user considerable costs in unnecessary analysis and testing.

The ideas discussed in this article are specific to software used in the design or manufacture of medical devices. With little change, the same

concepts could apply to software-controlled subsystems that are embedded in the medical device itself or to software modules that are included as part of the medical device embedded software. The main point is that MDMs are legally obligated to validate any software used for regulated purposes. Anything suppliers can do to assist the MDMs will become a marketing advantage for the supplier.

Outside the medical device industry, the needs are the same, although they might not be federally regulated. Instrument suppliers can turn their good engineering processes into a marketing advantage by packaging validation as an identifiable benefit to the customer.

## ABOUT THE AUTHOR:



David Vogel is the founder and president of Intertech Engineering Associates, Inc.

Dr. Vogel was a participant in a joint AAMI/FDA workgroup to develop a standard for Critical Device Software Validation which was subsequently included in the IEC 62304 Software

Lifecycle Standard. He was also a participant on the joint AAMI/FDA workgroup to develop a Technical Information Report (TIR) for Medical Device Software Risk Management. Currently, Dr. Vogel is a member of the AAMI/FDA workgroup developing a TIR on Quality System Software Validation.

A frequent lecturer for workshops and seminars on topics related to medical device development and validation, Dr. Vogel also is the author of numerous publications and holds several patents.

Dr. Vogel received a bachelor's degree in electrical engineering from Massachusetts Institute of Technology. He earned a master's degree in biomedical engineering, a master's degree in electrical and computer engineering, and a doctorate in biomedical engineering from the University of Michigan.

### *Intertech Service Offerings:*

*Risk Analysis and Management  
Software Design and Development  
Electronic Design and Development  
Requirements Development and Management  
Documentation and Traceability  
Verification and Validation  
Evaluations, Reviews, Inspections  
Planning  
Project Management  
Compliance Consulting and Training  
Manufacturing and Quality System Software Validation*

*Leverage INTERTECH's expertise to:*

*Reduce Project Risk*

*Shorten Time to Market*

*Cut Development and Test Cost*

*Assure Quality Products*

## ABOUT INTERTECH:

**Intertech Engineering Associates** has been helping medical device manufacturers bring their products to market since 1982. Through a distinct top-down service model, Intertech offers high-level consulting and hands-on engineering. By balancing technical expertise and practical business experience, we support clients through all phases of product development. While we do make your job easier, Intertech exists not to replace but to partner with clients to help balance the concerns of quality, time and cost.

With considerable experience in FDA regulatory compliance, our time-tested development process can anticipate and solve problems inexpensively on the planning board rather than through costly solutions later in the development, test, or post-deployment phases. By using deliberate processes, Intertech ensures an improvement in quality and can build client expertise.

**Call us today for more information or a free consultation at 781.801.1100**

**INTERTECH** Engineering Associates, Inc.

100 Lowder Brook Drive Suite 2500 Westwood, MA 02090 USA

www.inea.com - info@inea.com - Tel: (781) 801-1100 - Fax: (781) 801-1108